

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

JAN 12 2010

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:	
Name <u>Kyle Carrington</u>	ID# <u>5336</u> Date <u>1-14-2010</u>
Agency <u>UAFPD</u>	Phone # <u>474-7721</u>
Instrument Location <u>612 Yukon Drive Fairbanks, AK 99775</u>	
DataMaster cdm S/N <u>130301</u>	
Alco S/N <u>X 301583</u> Target Value <u>.084</u> High Pressure <u>800 PSI</u>	
Alco Test Values <u>.086</u> <u>.086</u> 1 st Alco 2 nd Alco	
Signature <u>[Signature]</u> <u>08</u> <u>1/22/10</u>	
(OVER)	

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz

Scientific Director

State Breath Alcohol Program

Subscribed and sworn before me this 20th day of January, 2010.

[Signature] (Notary Seal Stamp)

Carolyn M. Noland

Notary Public, State of Alaska

Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130301

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130301

JANUARY 14, 2010

OPERATOR'S NAME: KIC
CARRINGTON, KYLE
OPERATOR'S NUMBER: 5336
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI:

I
O.L. #: 0

DEPT/AGENCY: FBK7

CASE/REPORT: 0

TEST TYPE: U

ALCO TARGET VALUE: .004

ALCO S/N: X301583

--- BREATH ANALYSIS ---

.004 ADJUSTED FOR 29.38 in		
ALCO TARGET	.002	07:24
BLANK TEST	.000	07:25
INTERNAL STANDARD	VERIFIED	07:25
ALCO TU 29.38 in	.006	07:26
BLANK TEST	.000	07:27
SUBJECT SAMPLE	.000	07:27
BLANK TEST	.000	07:28
ALCO TU 29.38 in	.006	07:28
BLANK TEST	.000	07:29

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130301

JANUARY 14, 2010
TIME 07:22

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	40c
BAROMETER:	29.38 in
FLOW DETECTOR:	OKAY
PUMP	
RICH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=?@ABCDEF6
HIJKLMNPOQRSTUVWXYZ[\]^_`abcdefgijklmno
pqrstuvwxyz{|}~■